

Certificate



Certificate no.

T 09771006 14

License Holder:

Ault Incorporated
7105 Northland Terrace
Minneapolis MN 55428-1534
USA

Manufacturing Plant:

Ault Korea Corporation
Hwa-Seong-si
26-6 Wau-Ri, BongDam-eup
Kyunggi-Do
South Korea

Test report no.: USA-JAK 09771025 013

Client Reference: Tim Cassidy

Tested to: EN 60601-1:1990+A1+A2+A13

Certified Product: Medical Battery Charger

License Fee - Units

Addition:

Model Designation: BHW12150XXXXX1 (X=0-9, X1=A-Z)

Rated Voltage: AC 100-240V, 50-60Hz

Rated Current: 0.80-0.35A

Protection Class: I

Output Ratings DC: 14.7V/1.5A

Special Remarks: To be installed according to the licensee's installation instructions. EMC has to be investigated in the end user product according to EN 60601-1-2. X and X1 are not safety-relevant.

Appendix: 1 Pg. 1-3

Licensed Test mark:



Signatures

Stephan Schmitt
President

Dipl.-Ing. M. Raap
QA Certification Officer

Date of Issue

(day/mo/yr)

21/06/2004

Certificate



Certificate no.

T 09771006 15

License Holder:

Ault Incorporated
7105 Northland Terrace
Minneapolis MN 55428-1534
USA

Manufacturing Plant:

Ault Korea Corporation
Hwa-Seong-si
26-6 Wau-Ri, BongDam-eup
Kyunggi-Do
South Korea

Test report no.: USA-JAK 09771025 013

Client Reference: Tim Cassidy

Tested to: EN 60601-1:1990+A1+A2+A13

Certified Product: Medical Battery Charger

License Fee - Units


Change:

Following Manufacturing Plant no longer valid:

Ault Incorporated
7105 Northland Terrace
Minneapolis, MN 55428-1534

(K710355)

Licensed Test mark:



Signatures

Stephan Schmitt
President

Dipl.-Ing. M. Raap
QA Certification Officer

Date of Issue
(day/mo/yr)

21/06/2004