

# Certificate



Certificate no.

T 09771006 14

**License Holder:**

Ault Incorporated  
7105 Northland Terrace  
Minneapolis MN 55428-1534  
USA

**Manufacturing Plant:**

Ault Korea Corporation  
Hwa-Seong-si  
26-6 Wau-Ri, BongDam-eup  
Kyunggi-Do  
South Korea

**Test report no.:** USA-JAK 09771025 013

**Client Reference:** Tim Cassidy

**Tested to:** EN 60601-1:1990+A1+A2+A13

**Certified Product:** Medical Battery Charger

**License Fee - Units**

**Addition:**

**Model Designation:** BHW12150XXXXX1 (X=0-9, X1=A-Z)

**Rated Voltage:** AC 100-240V, 50-60Hz

**Rated Current:** 0.80-0.35A

**Protection Class:** I

**Output Ratings DC:** 14.7V/1.5A

**Special Remarks:** To be installed according to the licensee's installation instructions. EMC has to be investigated in the end user product according to EN 60601-1-2. X and X1 are not safety-relevant.

Appendix: 1 Pg. 1-3

**Licensed Test mark:**



**Signatures**

Stephan Schmitt  
President

Dipl.-Ing. M. Raap  
QA Certification Officer

**Date of Issue**

(day/mo/yr)

21/06/2004

# Certificate



Certificate no.

T 09771006 15

**License Holder:**

Ault Incorporated  
7105 Northland Terrace  
Minneapolis MN 55428-1534  
USA

**Manufacturing Plant:**

Ault Korea Corporation  
Hwa-Seong-si  
26-6 Wau-Ri, BongDam-eup  
Kyunggi-Do  
South Korea

**Test report no.:** USA-JAK 09771025 013

**Client Reference:** Tim Cassidy

**Tested to:** EN 60601-1:1990+A1+A2+A13

**Certified Product:** Medical Battery Charger

**License Fee - Units**


Change:

Following Manufacturing Plant no longer valid:

Ault Incorporated  
7105 Northland Terrace  
Minneapolis, MN 55428-1534

(K710355)

**Licensed Test mark:**



**Signatures**

Stephan Schmitt  
President

Dipl.-Ing. M. Raap  
QA Certification Officer

**Date of Issue**  
(day/mo/yr)

21/06/2004