

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE CERTIFICATS D'ESSAIS DES EQUIPEMENTS ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE

CERTIFICAT D'ESSAI OC

Product Produit	Medical Power Supply
Name and address of the applicant Nom et adresse du demandeur	BRIDGEPOWER CORP 964 GOSAEK-DONG GWONSEON-GU SUWON-SI GYEONGGI-DO 441-813 KOREA
Name and address of the manufacturer Nom et adresse du fabricant	BRIDGEPOWER CORP 964 GOSAEK-DONG GWONSEON-GU SUWON-SI GYEONGGI-DO 441-813 KOREA
Name and address of the factory Nom et adresse de l'usine	See Page 2
Note: When more than one factory, please report on page 2 Note: Lorsque il y plus d'une usine, veuillez utiliser la 2 ^{eme} page Ratings and principal characteristics Valeurs nominales et caractéristiques principales	See Page 2
Trademark (if any) Marque de fabrique (si elle existe)	None
Type of Manufacturer's Testing Laboratories used Type de programme du laboratoire d'essais constructeur Model / Type Ref. Ref. De type	(a)ENB1020(b)'(c)(d)(e)(f), BP(a)020(b)(c)(e)(f), See Page 2
Additional information (if necessary may also be reported on page 2) Les informations complémentaires (si nécessaire,, peuvent être indiqués sur la 2 ^{eme} page	Also investigated to EN 60601-1:2006
A sample of the product was tested and found to be in conformity with Un échantillon de ce produit a été essayé et a été considéré conforme à la	IEC 60601-1(ed.3)
As shown in the Test Report Ref. No. which forms part of this Certificate Comme indiqué dans le Rapport d'essais numéro de référence qui constitue partie de ce Certificat	E302267-20110519


This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



Date: 2011-05-24

UL International Demko A/S, Lyskaer 8, P.O. Box 514, DK-2730 Herlev, Denmark, Tel. +45 44 85 65 65, info.dk@dk.ul.com
www.ul-europe.com

Signature:



Jan-Erik Storgaard



Ref. Certif. No.

DK-3154

Model Details:

- (a) ENB1020(b)(c)(d)(e)(f), BP(a)020(b)(c)(e)(f);
- (a) can be A to Z for family related designs.
- (b) can be S for single output in model BP(a)020 series and (b)' can be A to Z for design revision changes in model (a)ENB1020 series.
- (c) can be 03 for 3.3Vdc, 05 for 5.0Vdc, 06 for 6.0Vdc, 07 for 7.5Vdc, 09 for 9.0Vdc, 12 for 12Vdc, 14 for 14Vdc, 15 for 15Vdc, 18 for 18Vdc, 24 for 24Vdc or 48 for 48Vdc output voltage.
- (d) can be 00 thru 99 for standards output cord options ("d)" is not provided in model BP(a)020series).
- (e) can be F or N or Q or B or H or G or M or C for input plug type.
- (f) can be 00 thru 99 for customer options, not related safety concerns.

Factories:

BRIDGEPOWER CORP
964 GOSAEK-DONG GWONSEON-GU
SUWON-SI GYEONGGI-DO 441-813 KOREA

WENDENG JEIL ELECTRONICS CO LTD
DONG SHOU GUANGZHOU LU KAIFA-QU
WENDENG-SHI SHANDONG CHINA

Ratings:

Rated Input; 100-240 Vac, 50-60 Hz, 0.5 A

Rated Output;

3.3Vdc, 3.0A or 5Vdc, 3.0A or 6Vdc, 2.5A or 7.5Vdc, 2.0A or 9Vdc, 2.0A or 12Vdc, 1.5A or 14Vdc, 1.2A or 15Vdc, 1.2A or 18Vdc, 1.0A or 24Vdc, 0.75 or 48Vdc, 0.4A or 48Vdc, 0.35A.

(Rated output voltage is designated in the model name designation system).

Additional information (if necessary)

Information complémentaire (si nécessaire)



UL International Demko A/S, Lyskaer 8, P.O. Box 514, DK-2730
Herlev, Denmark, Tel. +45 44 85 65 65, info.dk@dk.ul.com
www.ul-europe.com

Date: 2011-05-24

Signature:

Jan-Erik Storgaard