



Ref. Certif. No.

DK-22050

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE CERTIFICATS
D'ESSAIS DES EQUIPEMENTS ELECTRIQUES (IECEE)
METHODE OC

CB TEST CERTIFICATE
CERTIFICAT D'ESSAI OC

Product
Produit

AC-DC Adaptor

Name and address of the applicant
Nom et adresse du demandeur

BRIDGEPOWER CORP
964 GOSAEK-DONG
GWONSEON-GU
SUWON-SI GYEONGGI-DO 441-813 KOREA
BRIDGEPOWER CORP
964 GOSAEK-DONG
GWONSEON-GU
SUWON-SI GYEONGGI-DO 441-813 KOREA
BRIDGEPOWER CORP
964 GOSAEK-DONG
GWONSEON-GU
SUWON-SI GYEONGGI-DO 441-813 KOREA

Name and address of the manufacturer
Nom et adresse du fabricant

Name and address of the factory
Nom et adresse de l'usine

Note: When more than one factory, please report on page 2
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2^{ème} page

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

100-240Vac, 50-60Hz, 1.2 A.,

Trademark (if any)
Marque de fabrique (si elle existe)

NONE

Model / Type Ref.
Ref. De type

See appendix

Additional information (if necessary may also be reported
on page 2)
Les informations complémentaires (si nécessaire, peuvent
être indiqués sur la 2^{ème} page)

See appendix

A sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à la

PUBLICATION | EDITION
IEC 60601-1:1988 + A1:91+ A2:95 | 2nd

As shown in the Test Report Ref. No. which forms part of
this Certificate
Comme indiqué dans le Rapport d'essais numéro de
référence qui constitue partie de ce Certificat

E302267-A33-CB-1 dated 2011-03-10

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme National de Certification



Date:

2011-03-10

UL International Demko A/S

Lyskaer 8, P.O. Box 514, DK-2730 Herlev, Denmark
Tel: +45 44856565 Fax: +45 44856500

Signature

Jan-Erik Storgaard
Certification manager



Ref. Certif. No.

DK-22050

Model/Type reference:

BP(a)040(b)(c)(d)(e)

(a) can be A to Z for family related designs.

(b) can be S for single.

(c) can be 05, 09, 12, 15, 18, 24 or 48 for output voltage.

(d) can be F or N or Q for input plug type.

(e) can be 00 thru 99 for customer options, not related safety concerns.

(a)ENB1040(b)(c)(d)(e)(f)

(a) can be A to Z for family related designs.

(b) can be A to Z for design revision changes.

(c) can be 05, 09, 12, 15, 18, 24 or 48 for output voltage.

(d) can be 00 thru 99 for standard output cord options.

(e) can be F or N or Q for input plug type.

(f) can be 00 thru 99 for customer options, not related safety concerns.

Rating:

Rated Input; 100-240Vac, 50-60Hz, 1.2 A.

Rated Output; 5Vdc, 5.0A or 9Vdc, 4.0A or 12Vdc, 3.4A or 15Vdc, 2.7A or 18Vdc, 2.2A or 24Vdc, 1.7A or 48Vdc, 0.83A. Rated output voltage is designated in the model name designation system).

Additional information (if necessary)

Information complémentaire (si nécessaire)



**Underwriters
Laboratories**

Date:

2011-03-10

UL International Demko A/S

Lyskaer 8, P.O. Box 514, DK-2730 Herlev, Denmark

Tel: +45 44856565, Fax: +45 44856500

Signature:

Jan-Erik Storgaard
Certification manager