COVER PAGE FOR TEST REPORT

Product Category:	Power Supplies, Medical and Dental
Product Category CCN:	QQHM2, QQHM8
Test Procedure:	Component Recognition
Product:	AC-DC Adaptor
Model/Type Reference:	J(1)W180(2)(3)(4)(5)F(6) and (1)ENB1080(3)(4)(5)F(6) (1) may be C, M or P according to model name designation without any other difference. (2) may be J, K or C according to manufacturing site (J-JEC Korea, K-Ault Korea, C-Wendeng Jeil Electronics). (3) may be A to Z according to design revision. (4) may be 12, 15, 18 or 24 according to output voltage. (5) may be 00 to 99 according to the shape of output connector. (6) may be 00 to 99 or AA to ZZ according to customer's option.
Rating(s):	Rated Input: 100-240 Vac, 50-60Hz, 1.5A. Rated Output: +12V/6.5A, +15V/5.2A, +18V/4.33A or +24V/3.25A.
Standards:	UL 60601-1, 1st Edition, 2006-04-26 (Medical Electrical Equipment, Part 1: General Requirements for Safety) CAN/CSA-C22.2 No. 601.1-M90, 2005 (Medical Electrical Equipment - Part 1: General Requirements for Safety)
Applicant Name and Address:	JEC KOREA CORP 26-6 WAU-RI BONGDAM-EUP HWASEONG-SHI KYUNGGI-DO 445-897 KOREA

This Report includes the following parts, in addition to this cover page:

1. Specific Inspection Criteria
2. Specific Technical Criteria
3. Clause Verdicts
4. Critical Components
5. Test Results
6. National Differences
7. Enclosures

Copyright © 2008

This is to certify that representative samples of the products covered by this Test Report have been investigated in accordance with the above referenced Standards. The products have been found to comply with the requirements covering the category and the products are judged to be eligible for Follow-Up Service under the indicated Test Procedure. The manufacturer is authorized to use the UL Mark on such products which comply with this Test Report and any other applicable requirements of Underwriters Laboratories Inc. ('UL') in accordance with the Follow-Up Service Agreement. Only those products which properly bear the UL Mark are considered as being covered by UL's Follow-Up Service under the indicated Test Procedure.

The applicant is authorized to reproduce the referenced Test Report provided it is reproduced in its entirety.

UL authorizes the applicant to reproduce the latest pages of the referenced Test Report consisting of the first page of the Specific Technical Criteria through to the end of the Conditions of Acceptability.

Any information and documentation involving UL Mark services are provided on behalf of Underwriters Laboratories Inc. (UL) or any authorized licensee of UL.

Test Report By:

DongYoul Kim
Project Engineer
UL Korea, Ltd.

Reviewed By:

C. Y. Won

ChanYo Won
Section Manager
UL Korea, Ltd.

Copyright © 2008