

IEC SYSTEM FOR CONFORMITY TESTING AND
CERTIFICATION OF ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

SYSTEME CEI D'ESSAIS DE CONFORMITE ET DE CERTIFICATION
DES EQUIPEMENTS ELECTRIQUES (IECEE)
METHODE OC

CB TEST CERTIFICATE

CERTIFICAT D'ESSAI OC

Product
Produit

Switching Power Supply

Name and address of the applicant
Nom et adresse du demandeur

SL POWER ELECTRONICS CORP
BLDG A 6050 KING ST
VENTURA CA 93003, USA

Name and address of the manufacturer
Nom et adresse du fabricant

SL POWER ELECTRONICS CORP
BLDG A 6050 KING ST
VENTURA CA 93003, USA

Name and address of the factory
Nom et adresse de l'usine

For factories see information on page 2

Rating and principal characteristics
Valeurs nominales et caractéristiques principales

Input: 100-240 V~, 50-60 Hz, 3.0 A
For output ratings see information on page 2

Trademark (if any)
Marque de fabrique (si elle existe)



Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais constructeur

SMT

Model / Type Ref.
Ref. de type

MENT1150AWWXXYZ Series;
Where A is an alpha character A-Z, WW is a number 0-48, XX is a 2 digit number 00-99, Y is an alpha character A-Z, and Z is a number 00-99.

Additional information (if necessary)
Information complémentaire (si nécessaire)

This report comprises 2 enclosures.

A sample of the product was tested and found
to be in conformity with
*Un échantillon de ce produit a été essayé et a été
considéré conforme à la*

**IEC 60601-1 (1988) Second Edition,
with Amendment No. 1 (1991) and No. 2 (1995)** with the exception of:
Clause 36, Electromagnetic Compatibility, Clause 48, Biocompatibility and
Clause 52.1, Programmable Electronic Systems. Inclusive of CENELEC Common
Modifications. See Test Report for National Differences.

as shown in the Test Report Ref. No.
which forms part of this Certificate
*comme indiqué dans le Rapport d'essais numéro
de référence qui constitue partie de ce Certificat*

E116994-A60-CB-1

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme National de Certification



**Underwriters
Laboratories**

Underwriters Laboratories Inc. / GMA Certification Department, US
333 Pfingsten Road, Northbrook, IL 60062-2096
United States of America
TEL INT* +1 847 664 3008, FAX INT* +1 847 313 3008
email: jolanta.m.wroblewska@us.ul.com

Date: Issued: 2011 February 9

Signature:

Jolanta M. Wroblewska

IEC SYSTEM FOR CONFORMITY TESTING AND
CERTIFICATION OF ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

SYSTEME CEI D'ESSAIS DE CONFORMITE ET DE CERTIFICATION
DES EQUIPEMENTS ELECTRIQUES (IECEE)
METHODE OC

CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

1. INDUSTRIAS S L S A DE C V
COSTA RICA #60
COL CUAHUTEMOC MEXICALI BC, MEXICO
2. SL POWER ELECTRONICS XIANGHE
ANPING ECONOMIC & TECH DEVELOPING ZONE
XIANGHE HEBEI 065402, CHINA

Output: 48 V maximum,
11.67 A maximum,
or see below for standard output models.
MENT1150V12XXFZ: 12 V dc/11.67 A;
MENT1150V15XXFZ: 15 V dc/9.33 A;
MENT1150V18XXFZ: 18 V dc/7.78 A;
MENT1150V24XXFZ: 24 V dc/6.25 A;
MENT1150V28XXFZ: 28 V dc/5.36 A;
MENT1150V48XXFZ: 48 V dc/3.13 A.

Additional information (if necessary)
Information complémentaire (si nécessaire)

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme National de Certification




**Underwriters
Laboratories**

Date: Issued: 2011 February 9

Underwriters Laboratories Inc. / GMA Certification Department, US
333 Pfingsten Road, Northbrook, IL 60062-2096
United States of America
TEL INT* +1 847 664 3008 FAX INT* +1 847 313 3008
email: jolanta.m.wroblewska@us.ul.com

Signature:



Jolanta M. Wroblewska