

IEC SYSTEM FOR CONFORMITY TESTING AND
CERTIFICATION OF ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

SYSTEME CEI D'ESSAIS DE CONFORMITE ET DE CERTIFICATION
DES EQUIPEMENTS ELECTRIQUES (IECEE)
METHODE OC

CB TEST CERTIFICATE

CERTIFICAT D'ESSAI OC

Product <i>Produit</i>	Switching Power Supply
Name and address of the applicant <i>Nom et adresse du demandeur</i>	SL POWER ELECTRONICS CORP BLDG A 6050 KING ST VENTURA CA 93003, USA
Name and address of the manufacturer <i>Nom et adresse du fabricant</i>	SL POWER ELECTRONICS CORP BLDG A 6050 KING ST VENTURA CA 93003, USA
Name and address of the factory <i>Nom et adresse de l'usine</i>	For factories see information on page 2
Rating and principal characteristics <i>Valeurs nominales et caractéristiques principales</i>	Input: 100-240 V~, 50-60 Hz, 3.0 A For output ratings see information on page 2
Trademark (if any) <i>Marque de fabrique (si elle existe)</i>	 SMT
Type of Manufacturer's Testing Laboratories used <i>Type de programme du laboratoire d'essais constructeur</i>	
Model / Type Ref. <i>Ref. de type</i>	MENT1150AWWXXYZ Series; Where A is an alpha character A-Z, WW is a number 0-48, XX is a 2 digit number 00-99, Y is an alpha character A-Z, and Z is a number 00-99.
Additional information (if necessary) <i>Information complémentaire (si nécessaire)</i>	This report comprises 2 enclosures.
A sample of the product was tested and found to be in conformity with <i>Un échantillon de ce produit a été essayé et a été considéré conforme à la</i>	IEC 60601-1 (1988) Second Edition, with Amendment No. 1 (1991) and No. 2 (1995) with the exception of: Clause 36, Electromagnetic Compatibility, Clause 48, Biocompatibility and Clause 52.1, Programmable Electronic Systems. Inclusive of CENELEC Common Modifications. See Test Report for National Differences.
as shown in the Test Report Ref. No. which forms part of this Certificate <i>comme indiqué dans le Rapport d'essais numéro de référence qui constitue partie de ce Certificat</i>	E116994-A60-CB-1

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme National de Certification



**Underwriters
Laboratories**

Underwriters Laboratories Inc. / GMA Certification Department, US
333 Pfingsten Road, Northbrook, IL 60062-2096
United States of America
TEL INT* +1 847 664 3008, FAX INT* +1 847 313 3008
email: jolanta.m.wroblewska@us.ul.com

Date: Issued: 2011 February 9

Signature:

Jolanta M. Wroblewska

IEC SYSTEM FOR CONFORMITY TESTING AND
CERTIFICATION OF ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

SYSTEME CEI D'ESSAIS DE CONFORMITE ET DE CERTIFICATION
DES EQUIPEMENTS ELECTRIQUES (IECEE)
METHODE OC

CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

1. INDUSTRIAS S L S A DE C V
COSTA RICA #60
COL CUAHUTEMOC MEXICALI BC, MEXICO
2. SL POWER ELECTRONICS XIANGHE
ANPING ECONOMIC & TECH DEVELOPING ZONE
XIANGHE HEBEI 065402, CHINA

Output: 48 V maximum,
11.67 A maximum,
or see below for standard output models.
MENT1150V12XXFZ: 12 V dc/11.67 A;
MENT1150V15XXFZ: 15 V dc/9.33 A;
MENT1150V18XXFZ: 18 V dc/7.78 A;
MENT1150V24XXFZ: 24 V dc/6.25 A;
MENT1150V28XXFZ: 28 V dc/5.36 A;
MENT1150V48XXFZ: 48 V dc/3.13 A.

Additional information (if necessary)
Information complémentaire (si nécessaire)

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme National de Certification




**Underwriters
Laboratories**

Date: Issued: 2011 February 9

Underwriters Laboratories Inc. / GMA Certification Department, US
333 Pfingsten Road, Northbrook, IL 60062-2096
United States of America
TEL INT* +1 847 664 3008 FAX INT* +1 847 313 3008
email: jolanta.m.wroblewska@us.ul.com

Signature:



Jolanta M. Wroblewska