

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE**CERTIFICAT D'ESSAI OC**

Product
Produit

Medical Switching Power Supply, AC-DC Power Supply

Name and address of the applicant
Nom et adresse du demandeur

SL POWER ELECTRONICS CORP
BLDG A
6050 KING DR
VENTURA, CA 93003 USA

Name and address of the manufacturer
Nom et adresse du fabricant

SL POWER ELECTRONICS CORP
BLDG A
6050 KING DR
VENTURA, CA 93003 USA

Name and address of the factory
Nom et adresse de l'usine

INDUSTRIAS S L S A DE C V
CIRCUITO SIGLO XXI 2055
COL PARQUE INDUSTRIAL EX-XXI
MEXICALI, BC 21254 Mexico

Note: When more than one factory, please report on page 2
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2^{ème} page

Additional Information on page 2
See Page 2

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

Trademark (if any)
Marque de fabrique (si elle existe)



CTF Stage 3

Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais
constructeur

MC425D2412EF, MU425S18EXX, TU425S18EXX
See Page 2

Model / Type Ref.
Ref. De type

Additional Information on page 2

Additional information (if necessary may also be
reported on page 2)
Les informations complémentaires (si nécessaire,,
peuvent être indiqués sur la 2^{ème} page

A sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à la

IEC 60601-1(ed.3)

As shown in the Test Report Ref. No. which forms
part of this Certificate
Comme indiqué dans le Rapport d'essais numéro de
référence qui constitue partie de ce Certificat

E116994-A103-CB-2 issued on 2015-08-13

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2015-08-14

Signature:

Jolanta M. Wroblewska



Ref. Certif. No.

US-25795-UL

Model Details:

MU425S18EXX,TU425S18EXX (XX=00-99, or blank)

Factories:

SL XIANGHE POWER ELECTRONICS CORP
No.B-02-03,North side of Landscape Ave, Qibu District,
Environmental Industrial Park
Xianghe, 065400 Hebei China

Ratings:

Input: 100-240V~, 50-60Hz, 4.5A

Output:

For convection:

Main output: 18Vdc/14.67A max

Fan output: 12Vdc/0.5A max

Standby output: 5Vdc/2.0A max

For 200 LFM airflow:

Main output: 18Vdc/21.45A max

Fan output: 12Vdc/1.0A max

Standby output: 5Vdc/2.0A max

MC425D2412EF: 400 W max total power with 15 cfm integral fan

Main output: 24Vdc/15.7A max

Standby output: 5Vdc/2.0A max

Main output 2: 12 Vdc/8.0A max

Fan output: 12Vdc/1.0A max

Additional Information:

Additionally evaluated to EN 60601-1:2006/A11:2011/A12:2014; National Differences specified in the CB Test Report.
The risk management requirements of the standard were not addressed.

Additional information (if necessary)

Information complémentaire (si nécessaire)



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA



UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK



UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN



UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2015-08-14

Signature:

Jolanta M. Wroblewska