

## **COVER PAGE FOR TEST REPORT**

Product Category:	Power Supplies, Medical and Dental
Product Category CCN:	QQHM2, QQHM8
Test Procedure:	Component Recognition
Product:	AC-DC Adaptor
Model/Type Reference:	JMW140(a)(b)(c)(d)(e)(f), MW153(a)(b)(c)(d)(e)(f). (a) can be J, K or C for manufacturing site (J-Jec Korea, K-Ault Korea, C-Wendeng Jeil Electronics). (b) can be A to Z for design revision changes. (c) can be 05, 09, 12, 15, 18, 24 for output voltage. (d) can be 00 to 99 for the design of output metal connector. (e) can be F or N for input plug type (F-Class I type, N-Class II type). (f) can be 00 to 99 for customer options. See Enclosure-Miscellaneous for the nomenclature of model name in details.
Rating(s):	Rated Input: 100-240 Vac, 50-60Hz, 1.2 A. Rated Output: +5 Vdc/5 A or +9 Vdc/4A or +12 Vdc/3.4A or + 15Vdc/2.7A or +18 Vdc/2.2A or +24 Vdc/1.7A (Rated output voltage is designated in the model name designation system).
Standards:	UL 60601-1, First Edition (2003) CAN/CSA-C22.2 No.601.1-M90 with updates 1 and 2
Applicant Name and Address:	JEC KOREA CORP 26-6 WAU-RI BONGDAM-EUP HWASEONG-SHI KYUNGGI-DO 445-897 KOREA
This Report includes the following parts, in addition to this cover page:	
<ol style="list-style-type: none"><li>1. Specific Technical Criteria</li><li>2. Critical Components</li><li>3. Test Results</li><li>4. Enclosures</li></ol>	

Issue Date: 2006-11-20  
Amendment 1 2006-12-06

Page 2 of 2

Report Reference #

E302267-A4-UL-1

This is to certify that representative samples of the products covered by this Test Report have been investigated in accordance with the above referenced Standards. The products have been found to comply with the requirements covering the category and the products are judged to be eligible for Follow-Up Service under the indicated Test Procedure. The manufacturer is authorized to use the UL Mark on such products which comply with this Test Report and any other applicable requirements of Underwriters Laboratories Inc. ('UL') in accordance with the Follow-Up Service Agreement. Only those products which properly bear the UL Mark are considered as being covered by UL's Follow-Up Service under the indicated Test Procedure.

The applicant is authorized to reproduce the referenced Test Report provided it is reproduced in its entirety.

UL authorizes the applicant to reproduce the latest pages of the referenced Test Report consisting of the first page of the Specific Technical Criteria through to the end of the Conditions of Acceptability.


Any information and documentation provided to you involving UL Mark services are provided on behalf of Underwriters Laboratories Inc(ULI) or any authorized license of ULI.

Test Report By:



KiSuk Cho  
Project Engineer  
UL Korea, Ltd.

Reviewed By:



ChanYo Won  
Senior Project Engineer  
UL Korea, Ltd.