

COVER PAGE FOR TEST REPORT

Product Category:	Power Supplies, Medical and Dental
Product Category CCN:	QQHM2, QQHM8
Test Procedure:	Component Recognition
Product:	AC-DC Adaptor
Model/Type Reference:	JMW118(a)(b)(c)(d)(e)(f), MW172(a)(b)(c)(d)(e)(f), JMW110(a)(b)(c)(d)(e)(f), MW170(a)(b)(c)(d)(e)(f). (a) can be K or C for manufacturing site (K- Bridgepower Corp, C-Wendeng Jeil Electronics). (b) can be A to Z for design revision changes. (c) can be 03, 05, 06, 09, 12, 14, 15, 18, 24 for output voltage (03 is not provided in model JMW110 series and MW170 series). (d) can be 00 to 99 for the design of output metal connector. (e) can be F or N or B or H or G or M or C for input plug type. See Enclosure-Photographs for each plug-type configuration [F-Class I appliance inlet type, N-Class II appliance inlet type, B-Class II direct-plug-in for North America(NEMA 5-15P), China , Japan, Changeable Direct plugin type for Model MW series only, H- Class II direct-plug-in for Australia(AS/NZS 3112) , G-Class II direct-plug-in for British(BS 1364) , M-Class II direct-plug-in for European(CEE 7/16)]. C- Changeable Direct plug-in type for National markets for Model JMW series only.(f) can be 00 to 99 for customer options. See Enclosure-Miscellaneous for the nomenclature of model name in details.
Rating(s):	(For JMW118 and MW172 Series) Rated Input: 100-240 Vac, 50-60Hz, 0.5A. Rated Output: +3.3Vdc/3.0A or +5 Vdc/3.0A or +6Vdc/2.5A or +9.0 Vdc/2.0A or +12.0 Vdc/1.5A or +15.0Vdc/1.2A or +18.0Vdc/1.0A or +24.0Vdc/0.75A. (For JMW110 and MW170 Series) Rated Input: 100-240 Vac, 50-60Hz, 0.4A. Rated Output: +5 Vdc/2.0A or +6Vdc/1.67A or +9.0 Vdc/1.5A or +12.0 Vdc/0.6A or +12.0 Vdc/1.25A or +14.0Vdc/0.85A or +15.0Vdc/1.0A or +18.0Vdc/0.84A or +24.0Vdc/0.63A. (Rated output voltage is designated in the model name designation system).
Standards:	UL 60601-1, 1st Edition, 2006-04-26 (Medical Electrical Equipment, Part 1: General Requirements for Safety) CAN/CSA-C22.2 No. 601.1-M90, 2005 (Medical Electrical Equipment - Part 1: General Requirements for Safety)
Applicant Name and Address:	BRIDGEPOWER CORP 964 GOSAEK-DONG GWONSEON-GU SUWON-SI GYEONGGI-DO 441-813 KOREA

This Report includes the following parts, in addition to this cover page:

1. Specific Technical Criteria
2. Clause Verdicts
3. Critical Components
4. Enclosures

Issue Date: 2006-12-07
Amendment 3 2008-12-29

Page 3 of 3

Report Reference #

E302267-A12-UL-1


This is to certify that representative samples of the products covered by this Test Report have been investigated in accordance with the above referenced Standards. The products have been found to comply with the requirements covering the category and the products are judged to be eligible for Follow-Up Service under the indicated Test Procedure. The manufacturer is authorized to use the UL Mark on such products which comply with this Test Report and any other applicable requirements of Underwriters Laboratories Inc. ('UL') in accordance with the Follow-Up Service Agreement. Only those products which properly bear the UL Mark are considered as being covered by UL's Follow-Up Service under the indicated Test Procedure.

The applicant is authorized to reproduce the referenced Test Report provided it is reproduced in its entirety.

UL authorizes the applicant to reproduce the latest pages of the referenced Test Report consisting of the first page of the Specific Technical Criteria through to the end of the Conditions of Acceptability.


Any information and documentation involving UL Mark services are provided on behalf of Underwriters Laboratories Inc. (UL) or any authorized licensee of UL.

Test Report By:



DongHa Lee
Engineer
UL Korea, Ltd.

Reviewed By:



ChanYo Won
Section Manager
UL Korea, Ltd.