



Ref. Certif. No.

US-17893-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE CERTIFICATS D'ESSAIS DES EQUIPEMENTS ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE

CERTIFICAT D'ESSAI OC

Product
Produit

Component – Switching Power Supply

Name and address of the applicant
Nom et adresse du demandeur

SL POWER ELECTRONICS CORP
6050 KING DRVIE, BLDG. A
VENTURA, CA 93003, USA

Name and address of the manufacturer
Nom et adresse du fabricant

SL POWER ELECTRONICS CORP
6050 KING DRIVE, BLDG. A
VENTURA, CA 93003, USA

Name and address of the factory
Nom et adresse de l'usine

See Page 2

Note: When more than one factory, please report on page 2
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2^{ème} page

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

See Page 2

Trademark (if any)
Marque de fabrique (si elle existe)



Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais constructeur
Model / Type Ref.
Ref. De type

SMT

GSM11-XYZ-XXXG, GSM11-10PP-108G, GSM11-10PP-104G,
See Page 2

Additional information (if necessary may also be reported on page 2)
Les informations complémentaires (si nécessaire,, peuvent être indiqués sur la 2^{ème} page

This report comprises 2 enclosures.

A sample of the product was tested and found to be in conformity with
Un échantillon de ce produit a été essayé et a été considéré conforme à la

IEC 60601-1(ed.3)

As shown in the Test Report Ref. No. which forms part of this Certificate
Comme indiqué dans le Rapport d'essais numéro de référence qui constitue partie de ce Certificat

11CA43006 issued on 2011-10-12

This CB Test Certificate is issued by the National Certification Body

Ce Certificat d'essai OC est établi par l'Organisme National de Certification



Date: 2011-10-20

Underwriters Laboratories Inc. / GMA Certification Department, US
333 Pflingsten Road, Northbrook, IL 60062-2096
United States of America
TEL INT* +1 847 664 3008, FAX INT* +1 847 313 3008
email: jolanta.m.wroblewska@us.ul.com
Signature:

Jolanta M. Wroblewska



Ref. Certif. No.

US-17893-UL

Model Details:

GSM11-XYZ-XXXG;
where X represents the output voltage which may be any number from 3 thru 28; Y indicates the type of input connector used which may be the letters A, B, P or T; Z indicates the type of output connector used which may be the letters A through Z; Models may also be followed by -XXX, where -XXX is any number from 001 to 999, which is used to designate value added configurations that have no impact on safety.

Factories:

SL POWER ELECTRONICS XIANGHE
NO.4, SHUANGXING NORTH ROAD,
XIANGHE ECONOMIC AND TECHNOLOGICAL DEVELOPMENT ZONE,
XIANGHE COUNTY, HEBEI PROVINCE, CHINA

SL POWER ELECTRONICS MEXICALI, MX
S.A. DE C.V.
CIRCUITO SIGLO XXI # 2205
COLONIA: PARQUE INDUSTRIAL EX-XXI
CP. 21254, MEXICALI, B.C., MEXICO

Ratings:

Input: 100-240 V ac, 0.3 A, 50/60 Hz

Output: 2.2 A or 11 W maximum output.

GSM11-3: 3.3 Vdc, 2.2A; GSM11-5: 5.1 Vdc, 2.2A; GSM11-10PP-104G: 10 Vdc, 1.1A; GSM11-10PP-108G: 10 Vdc, 1.1A; GSM11-12: 12 Vdc, 0.92A; GSM11-15: 15 Vdc, 0.74A; GSM11-24: 24 Vdc, 0.46A; GSM11-28: 28 Vdc, 0.4 A

Additional Information:

Additionally evaluated to EN 60601-1 (2006) Second Edition.
See Test Report for National Differences.

**Additional information (if necessary)
Information complémentaire (si nécessaire)**



Date: 2011-10-20

Underwriters Laboratories Inc. / GMA Certification Department, US
333 Pfingsten Road, Northbrook, IL 60062-2096
United States of America
TEL INT* +1 847 664 3008, FAX INT* +1 847 313 3008
email: jolanta.m.wroblewska@us.ul.com

Signature:

Jolanta M. Wroblewska