

DEMKO CERTIFICATE

Certificate No. D-00626
Page 1/3
Date of Issue 2011-10-20

Certificate Holder SL POWER ELECTRONICS CORP
6050 KING DRVIE, BLDG. A
VENTURA, CA 93003, USA

Manufacturer SL POWER ELECTRONICS CORP
6050 KING DRVIE, BLDG. A
VENTURA, CA 93003, USA

Production site See Page 2

Certified Product Component – Switching Power Supply
Model GSM11-XYZ-XXXG, GSM11-10PP-108G, GSM11-10PP-104G, See Page 2
Trademark



Rated Voltage / Frequency See Page 2
Rated Current / Power See Page 2
Insulation Class See Page 2
Degree of protection (IP) See Page 2
Tested acc. to EN 60601-1:2006
Test Report No. 11CA43006 issued on 2011-10-12
Additional
Expire date 2021-10-20

Certification Manager
Jan-Erik Storgaard

Certification Body UL International Demko A/S, Borupvang 5A, DK-2750
Ballerup, Denmark, Tel. +45 44 85 65 65, info.dk@dk.ul.com
www.ul-europe.com

The product and production sites listed on the certificate comply with the D-mark requirements and the UL Global Service Agreement, with reference to Terms and Conditions for the D mark. The Owner of the certificate is entitled to use the d or for cables <DEMKO> for the products listed on the certificate and manufactured at the production sites listed. UL has to be informed in writing about any changes to the product or production site in accordance with the Term and Conditions of the D mark. The validity of the certificate is shortened if the EU legislation require re-testing and re-certification due to new standards or amendments coming into force before the expiry date.

Appendix DEMKO CERTIFICATE

Certificate No. D-00626
Page 2/3
Date of Issue 2011-10-20

Model Details:

GSM11-XYZ-XXXG;

where X represents the output voltage which may be any number from 3 thru 28; Y indicates the type of input connector used which may be the letters A, B, P or T; Z indicates the type of output connector used which may be the letters A through Z; Models may also be followed by -XXX, where -XXX is any number from 001 to 999, which is used to designate value added configurations that have no impact on safety.

Production Sites:

SL POWER ELECTRONICS XIANGHE
NO.4,SHUANGXING NORTH ROAD,
XIANGHE ECONOMIC AND TECHNOLOGICAL DEVELOPMENT ZONE,
XIANGHE COUNTY, HEBEI PROVINCE, CHINA

SL POWER ELECTRONICS MEXICALI, MX
S.A. DE C.V.
CIRCUITO SIGLO XXI # 2205
COLONIA: PARQUE INDUSTRIAL EX-XXI
CP. 21254, MEXICALI, B.C., MEXICO

Ratings:

Input: 100-240 V ac, 0.3 A, 50/60 Hz

Output: 2.2 A or 11 W maximum output.

GSM11-3: 3.3 Vdc, 2.2A; GSM11-5: 5.1 Vdc, 2.2A; GSM11-10PP-104G: 10 Vdc, 1.1A; GSM11-10PP-108G: 10 Vdc, 1.1A;
GSM11-12: 12 Vdc, 0.92A; GSM11-15: 15 Vdc, 0.74A; GSM11-24: 24 Vdc, 0.46A; GSM11-28: 28 Vdc, 0.4 A

Certification Body

UL International Demko A/S, Borupvang 5A, DK-2750
Ballerup, Denmark, Tel. +45 44 85 65 65, info.dk@dk.ul.com
www.ul-europe.com

Appendix DEMKO Certificate

Certification Mark	D-mark
Certificate No.	D-00626
Page	3/3
Date of Issue	2011-10-20



Certification Body

UL International Demko A/S, Borupvang 5A, DK-2750
Ballerup, Denmark, Tel. +45 44 85 65 65, info.dk@dk.ul.com
www.ul-europe.com



the standard in safety

Underwriters
Laboratories