



Ref. Certif. No.

US-17777-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE CERTIFICATS D'ESSAIS DES EQUIPEMENTS ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE

CERTIFICAT D'ESSAI OC

Product
Produit

Component – Switching Power Supply

Name and address of the applicant
Nom et adresse du demandeur

SL POWER ELECTRONICS CORP
6050 KING DRVIE BLDG. A
VENTURA CA 93003, USA

Name and address of the manufacturer
Nom et adresse du fabricant

SL POWER ELECTRONICS CORP
6050 KING DRVIE BLDG. A
VENTURA CA 93003, USA

Name and address of the factory
Nom et adresse de l'usine

See Page 2

Note: When more than one factory, please report on page 2
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2^{ème} page

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

Input: 100-240V~, 50/60Hz, 1.0A
Output: See Test Report for output ratings.

Trademark (if any)
Marque de fabrique (si elle existe)



Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais constructeur
Model / Type Ref.
Ref. De type

SMT

MINT1045VWWXXYZZ, See Page 2

Additional information (if necessary may also be reported on page 2)
Les informations complémentaires (si nécessaire,, peuvent être indiqués sur la 2^{ème} page

This CB Test Report comprises 2 enclosures.
See Test Report for National Differences.

A sample of the product was tested and found to be in conformity with
Un échantillon de ce produit a été essayé et a été considéré conforme à la

IEC 60601-1(ed.3)

As shown in the Test Report Ref. No. which forms part of this Certificate
Comme indiqué dans le Rapport d'essais numéro de référence qui constitue partie de ce Certificat

11CA05392 issued on 2011-09-27

This CB Test Certificate is issued by the National Certification Body

Ce Certificat d'essai OC est établi par l'Organisme National de Certification



Underwriters Laboratories Inc. / GMA Certification Department, US
333 Pfingsten Road, Northbrook, IL 60062-2096
United States of America
TEL INT* +1 847 664 3008, FAX INT* +1 847 313 3008
email: jolanta.m.wroblewska@us.ul.com

Signature:

Jolanta M. Wroblewska

Date: 2011-09-28



Ref. Certif. No.

US-17777-UL

Model Details:

MINT1045VWWXXYZZ,

Where V represents the generational differences which may be either A or B (A is for Class I and B is for Class II construction), WW represents the output voltage which may be any number from 12 to 48, XX represents output connector which may be any two alphanumeric digits, Y represents the input connector which may be any letter from A thru Z, and ZZ represents non-safety related customer options and RoHS statements which may be any two alphanumeric digits.

Factories:

SL XIANGHE POWER ELECTRONICS CORP
NO.4, SHUANGXING NORTH ROAD
XIANGHE ECONOMIC AND TECHNOLOGICAL DEVELOPMENT ZONE
XIANGHE COUNTY HEBEI PROVINCE, CHINA

Additional information (if necessary)
Information complémentaire (si nécessaire)



Date: 2011-09-28

Underwriters Laboratories Inc. / GMA Certification Department, US
333 Pfingsten Road, Northbrook, IL 60062-2096
United States of America
TEL INT* +1 847 664 3008, FAX INT* +1 847 313 3008
email: jolanta.m.wroblewska@us.ul.com

Signature:

Jolanta M. Wroblewska